

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1650

Name: RITTER 1  
Sterilizer ID: V2178198  
02/17/2022 01:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 175.0 F 3.0  
4:00 192.9 F 7.0  
6:00 215.2 F 12.1  
8:00 225.2 F 6.9  
10:00 237.1 F 12.0  
12:00 235.8 F 9.2  
14:00 238.5 F 9.5  
16:00 247.8 F 14.3  
18:00 260.3 F 21.1  
20:00 271.4 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.4 F 28.9  
2:00 271.5 F 28.6  
3:00 272.3 F 28.9  
4:00 271.4 F 28.4

Min 271.3 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 231.0 F 0.0  
10:00 199.9 F 0.0  
15:00 196.2 F 0.0  
20:00 198.2 F 0.0  
25:00 193.1 F 0.0  
30:00 194.8 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:43  
CYCLE COMPLETE

02/17/2022 02:56 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02