

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3749

Name: RITTER2  
Sterilizer ID: V2313228  
07/02/2024 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.7 F 0.0  
2:00 99.2 F 0.5  
4:00 104.9 F 1.9  
6:00 127.1 F 4.4  
8:00 154.0 F 7.3  
10:00 190.2 F 10.3  
12:00 214.8 F 6.4  
14:00 210.6 F 5.9  
16:00 221.5 F 9.3  
18:00 231.8 F 13.2  
20:00 232.3 F 9.2  
22:00 240.4 F 11.8  
24:00 241.9 F 12.1  
26:00 252.7 F 17.5  
28:00 262.5 F 23.1  
30:00 271.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.7 F 29.8  
2:00 271.4 F 29.3  
3:00 272.5 F 29.9  
4:00 271.6 F 29.3

Min 271.3 F 29.2  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
5:00 193.1 F 0.0  
10:00 197.6 F 0.0  
15:00 196.0 F 0.0  
20:00 197.3 F 0.0  
25:00 197.7 F 0.0  
30:00 199.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:40  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:09:19  
CYCLE COMPLETE

07/02/2024 12:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_