
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004293
8:57 06/09/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:34
STERILIZING 3:48
136.4 C 220kPa 3:48
136.3 C 224kPa 4:18
136.4 C 224kPa 4:48
136.3 C 222kPa 5:18
136.2 C 222kPa 5:48
136.2 C 220kPa 6:18
136.4 C 221kPa 6:48
136.4 C 222kPa 7:18
136.4 C 222kPa 7:18
VENTING 7:18
AIR DRYING 9:03
CYCLE COMPLETE 9:08

Digital Signature #
DA32C638E9577324
