

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003026  
10:51 26/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:09  
STERILIZING 3:58  
136.0 C 220kPa 3:58  
136.1 C 223kPa 4:28  
136.3 C 224kPa 4:58  
136.4 C 224kPa 5:28  
136.3 C 224kPa 5:58  
136.3 C 224kPa 6:28  
136.3 C 224kPa 6:58  
136.4 C 224kPa 7:28  
136.4 C 224kPa 7:28  
VENTING 7:29  
AIR DRYING 8:20  
CYCLE COMPLETE 8:55

-----  
Digital Signature #  
A8C16F05F83C12BC  
-----