
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000845
11:51 06/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:25
STERILIZING 4:47
136.4 C 223kPa 4:47
136.4 C 233kPa 5:17
136.2 C 231kPa 5:47
136.4 C 233kPa 6:17
136.4 C 232kPa 6:47
136.4 C 232kPa 7:17
136.4 C 233kPa 7:47
136.4 C 233kPa 8:17
136.4 C 233kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 11:20

Digital Signature #
B8E51340E828CF7E
