

-----  
Midmark M11 - v1.0.5  
Total Cycles: 372

Name: RITTER2  
Sterilizer ID: V2313228  
09/16/2021 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.8 F 0.0  
2:00 168.7 F 3.6  
4:00 191.1 F 8.8  
6:00 216.1 F 8.8  
8:00 224.5 F 8.7  
10:00 236.6 F 11.1  
12:00 239.4 F 12.0  
14:00 240.0 F 11.2  
16:00 253.4 F 17.9  
18:00 265.0 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.5 F 29.8  
2:00 271.8 F 29.4  
3:00 272.8 F 29.8  
4:00 271.8 F 29.2  
  
Min 271.7 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 194.1 F 0.0  
10:00 198.5 F 0.0  
15:00 197.8 F 0.0  
20:00 193.9 F 0.0  
25:00 194.7 F 0.0  
30:00 192.9 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

09/16/2021 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_