
Midmark M11 - v1.0.5
Total Cycles: 3020

Name: RITTER2
Sterilizer ID: V2313228
10/17/2023 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.1 F 0.0
2:00 169.8 F 3.8
4:00 192.6 F 8.5
6:00 212.8 F 8.7
8:00 223.7 F 9.4
10:00 234.5 F 8.7
12:00 244.5 F 13.5
14:00 244.5 F 12.9
16:00 258.1 F 20.2
18:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 273.1 F 29.7
2:00 272.0 F 29.9
3:00 272.4 F 29.6
4:00 272.1 F 29.9

Min 271.5 F 29.2
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 241.2 F 0.0
10:00 214.5 F 0.0
15:00 236.3 F 0.0
20:00 196.7 F 0.0
25:00 193.6 F 0.0
30:00 193.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

10/17/2023 02:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____