

-----  
Midmark M11 - v1.0.5  
Total Cycles: 63

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2021 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.3 F 0.0  
2:00 156.1 F 2.6  
4:00 172.5 F 6.2  
6:00 194.5 F 10.7  
8:00 212.4 F 5.1  
10:00 222.9 F 9.5  
12:00 231.4 F 9.7  
14:00 239.0 F 12.1  
16:00 240.7 F 11.1  
18:00 252.8 F 17.4  
20:00 263.7 F 23.9  
22:00 271.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 271.6 F 29.1  
2:00 271.9 F 29.2  
3:00 272.1 F 29.4  
4:00 272.1 F 29.6

Min 271.3 F 29.1  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
5:00 240.8 F 0.0  
10:00 221.3 F 0.0  
15:00 225.1 F 0.0  
20:00 189.3 F 0.0  
25:00 185.6 F 0.0  
30:00 183.2 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 22:02  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

06/01/2021 11:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_