

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000373  
10:58 21/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:38  
STERILIZING 3:39  
136.2 C 231kPa 3:39  
136.2 C 232kPa 4:09  
136.4 C 233kPa 4:39  
136.4 C 233kPa 5:09  
136.4 C 233kPa 5:39  
136.1 C 230kPa 6:09  
136.4 C 232kPa 6:39  
136.3 C 233kPa 7:09  
136.3 C 233kPa 7:09  
VENTING 7:09  
AIR DRYING 8:03  
CYCLE COMPLETE 8:13  
-----

Digital Signature #  
DF8661FBCB52EE18  
-----