

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007574  
8:35 19/06/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:07  
STERILIZING 4:01  
136.4 C 220kPa 4:01  
136.4 C 224kPa 4:31  
136.0 C 222kPa 5:01  
136.0 C 222kPa 5:31  
136.4 C 219kPa 6:01  
136.4 C 224kPa 6:31  
136.2 C 223kPa 7:01  
136.2 C 222kPa 7:31  
136.2 C 222kPa 7:31  
VENTING 7:31  
AIR DRYING 8:28  
CYCLE COMPLETE 11:10  
-----

Digital Signature #  
42CC394F856A1564  
-----