
Midmark M11 - v1.0.0
Total Cycles: 4488

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2024 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 168.1 F 3.3
4:00 177.0 F 7.0
6:00 210.6 F 11.5
8:00 221.9 F 5.4
10:00 231.0 F 10.0
12:00 235.6 F 8.5
14:00 240.6 F 11.6
16:00 238.5 F 9.8
18:00 251.4 F 16.0
20:00 262.3 F 22.3
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 273.1 F 29.3
2:00 271.5 F 28.7
3:00 272.8 F 29.4
4:00 271.8 F 28.7

Min 271.5 F 28.7
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 235.9 F 0.0
10:00 226.6 F 0.0
15:00 219.1 F 0.0
20:00 225.3 F 0.0
25:00 229.2 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:14
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

04/24/2024 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____