
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006432
13:44 10/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:02
STERILIZING 4:07
136.4 C 224kPa 4:07
136.4 C 224kPa 4:37
136.4 C 225kPa 5:07
136.4 C 225kPa 5:37
136.4 C 223kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 7:07
136.2 C 223kPa 7:37
136.2 C 223kPa 7:37
VENTING 7:37
AIR DRYING 8:35
CYCLE COMPLETE 9:05

Digital Signature #
E0BD1B242B114EEB
