

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4146

Name: RITTER2  
Sterilizer ID: V2313228  
11/21/2024 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 80.9 F 0.0  
4:00 80.8 F 0.0  
6:00 99.6 F 0.3  
8:00 135.0 F 0.4  
10:00 180.3 F 0.8  
12:00 195.4 F 2.9  
14:00 209.9 F 6.3  
16:00 221.8 F 10.2  
18:00 230.1 F 9.2  
20:00 234.7 F 9.6  
22:00 238.4 F 10.3  
24:00 242.9 F 12.4  
26:00 240.8 F 11.2  
28:00 253.2 F 17.4  
30:00 264.0 F 23.8  
32:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.5 F 29.3  
2:00 272.0 F 29.3  
3:00 272.0 F 29.0  
4:00 272.4 F 29.6  
  
Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 222.3 F 0.0  
10:00 186.6 F 0.0  
15:00 178.7 F 0.0  
20:00 173.8 F 0.0  
25:00 171.5 F 0.0  
30:00 170.0 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:03  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:13  
CYCLE COMPLETE

11/21/2024 11:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_