
Midmark M11 - v1.0.0
Total Cycles: 2609

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.2 F 0.0
2:00 204.9 F 3.4
4:00 198.2 F 8.7
6:00 221.9 F 8.1
8:00 226.9 F 9.9
10:00 232.8 F 8.0
12:00 243.8 F 13.3
14:00 243.0 F 12.1
16:00 256.4 F 19.0
18:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.7
3:00 272.5 F 29.2
4:00 271.5 F 28.5

Min 271.3 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 231.6 F 0.0
10:00 231.6 F 0.0
15:00 229.0 F 0.0
20:00 228.4 F 0.0
25:00 231.1 F 0.0
30:00 232.1 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

11/28/2022 10:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____