

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001199  
14:26 28/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:25  
PRESSURIZING 4:26  
STERILIZING 5:41  
136.0 C 220kPa 5:41  
136.2 C 223kPa 6:11  
136.4 C 226kPa 6:41  
136.4 C 223kPa 7:11  
136.4 C 226kPa 7:41  
136.4 C 226kPa 8:11  
136.3 C 224kPa 8:41  
136.3 C 225kPa 9:11  
136.3 C 225kPa 9:11  
VENTING 9:12  
AIR DRYING 10:08  
CYCLE COMPLETE 17:21  
-----

Digital Signature #  
D9AB1B3B06FC0010  
-----