
Midmark M11 - v1.0.0
Total Cycles: 1498

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2022 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 163.8 F 2.0
4:00 171.5 F 5.2
6:00 202.4 F 9.4
8:00 230.3 F 6.3
10:00 232.0 F 8.6
12:00 240.0 F 13.5
14:00 241.0 F 10.9
16:00 236.7 F 8.9
18:00 250.9 F 15.6
20:00 262.5 F 22.2
22:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 28.8
1:00 272.6 F 28.8
2:00 272.5 F 28.9
3:00 272.6 F 29.1
4:00 272.4 F 29.3

Min 271.7 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 217.0 F 0.0
10:00 226.1 F 0.0
15:00 226.4 F 0.0
20:00 226.8 F 0.0
25:00 228.6 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

01/05/2022 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____