
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002580
7:48 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:30
STERILIZING 3:32
136.0 C 220kPa 3:32
136.2 C 222kPa 4:02
136.4 C 224kPa 4:32
136.2 C 223kPa 5:02
136.4 C 224kPa 5:32
136.1 C 221kPa 6:02
136.4 C 224kPa 6:32
136.5 C 225kPa 7:02
136.5 C 225kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 8:17

Digital Signature #
AF1C6CAEC51F1861
