

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3368

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2023 01:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.6 F 0.0  
2:00 153.8 F 0.0  
4:00 173.9 F 2.5  
6:00 196.8 F 6.3  
8:00 210.1 F 10.5  
10:00 219.2 F 5.8  
12:00 230.3 F 9.3  
14:00 237.3 F 10.8  
16:00 241.2 F 11.6  
18:00 238.8 F 9.9  
20:00 251.8 F 16.2  
22:00 262.8 F 22.5  
24:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.6 F 28.2  
3:00 272.1 F 29.0  
4:00 272.4 F 28.8

Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
5:00 210.9 F 0.0  
10:00 214.7 F 0.0  
15:00 214.1 F 0.0  
20:00 214.8 F 0.0  
25:00 213.3 F 0.0  
30:00 209.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:02:48  
CYCLE COMPLETE

06/09/2023 02:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_