
Midmark M11 - v1.0.0
Total Cycles: 849

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 07:52 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 155.3 F 1.5
4:00 190.8 F 4.8
6:00 203.4 F 1.3
8:00 215.4 F 4.7
10:00 222.5 F 4.0
12:00 232.3 F 7.3
14:00 232.0 F 6.8
16:00 245.3 F 12.7

STERILIZING
mm:ss Degrees PSI
0:00 251.3 F 15.8
1:00 252.5 F 16.2
2:00 251.3 F 15.8
3:00 252.6 F 16.3
4:00 251.8 F 15.8
5:00 252.0 F 16.2
6:00 252.6 F 16.3
7:00 252.1 F 16.0
8:00 251.4 F 15.7
9:00 252.2 F 16.3
10:00 252.7 F 16.4
11:00 252.5 F 16.2
12:00 252.0 F 16.0
13:00 251.6 F 15.8
14:00 251.5 F 15.9
15:00 252.3 F 16.4
16:00 252.7 F 16.4
17:00 252.7 F 16.4
18:00 252.5 F 16.2
19:00 252.2 F 16.1
20:00 251.9 F 15.9
21:00 251.6 F 15.8
22:00 251.4 F 15.8
23:00 251.8 F 16.1
24:00 252.3 F 16.4
25:00 252.7 F 16.5
26:00 252.8 F 16.4
27:00 252.7 F 16.4
28:00 252.7 F 16.3
29:00 252.4 F 16.2
30:00 252.2 F 16.1

Min 251.3 F 15.7
Max 252.8 F 16.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 208.6 F 0.0
10:00 181.8 F 0.0
15:00 175.7 F 0.0
20:00 169.8 F 0.0
25:00 164.4 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 17:39

STERILIZING: 30:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 01:21:35
CYCLE COMPLETE

05/31/2021 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U01