
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000838
11:02 06/04/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:02
PRESSURIZING 6:59
STERILIZING 8:42
136.0 C 220kPa 8:42
136.2 C 222kPa 9:12
136.4 C 224kPa 9:42
135.8 C 218kPa 10:12
136.1 C 221kPa 10:42
136.4 C 223kPa 11:12
136.2 C 221kPa 11:42
136.4 C 223kPa 12:12
136.4 C 223kPa 12:12
VENTING 12:12
AIR DRYING 13:21
CYCLE COMPLETE 43:21

Digital Signature #
6BE0FBC71E8DF2F4
