
Midmark M11 - v1.0.5
Total Cycles: 2211

Name: RITTER2
Sterilizer ID: V2313228
04/03/2023 04:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 183.4 F 4.0
4:00 210.4 F 9.3
6:00 220.6 F 7.4
8:00 233.8 F 11.2
10:00 235.5 F 9.5
12:00 237.4 F 9.2
14:00 250.8 F 16.2
16:00 263.5 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.4
2:00 272.1 F 29.8
3:00 272.1 F 29.2
4:00 272.4 F 29.9

Min 271.4 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 227.4 F 0.0
10:00 220.8 F 0.0
15:00 218.8 F 0.0
20:00 209.6 F 0.0
25:00 216.9 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

04/03/2023 05:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____