
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002438
15:00 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:04
136.0 C 220kPa 3:04
136.3 C 225kPa 3:34
136.4 C 226kPa 4:04
136.1 C 223kPa 4:34
136.2 C 221kPa 5:04
136.4 C 226kPa 5:34
136.4 C 225kPa 6:04
136.4 C 225kPa 6:34
136.4 C 225kPa 6:34
VENTING 6:34
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
E59C0514A22763B6
