
Midmark M11 - v1.0.0
Total Cycles: 2202

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2022 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 162.6 F 2.9
4:00 172.4 F 6.8
6:00 193.0 F 11.3
8:00 216.2 F 5.1
10:00 226.1 F 10.3
12:00 230.5 F 7.5
14:00 240.5 F 12.1
16:00 238.5 F 10.0
18:00 251.2 F 16.4
20:00 262.8 F 22.9
22:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.2 F 29.2
2:00 271.3 F 28.7
3:00 272.2 F 29.3
4:00 271.4 F 28.6

Min 271.2 F 28.6
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 213.0 F 0.0
10:00 206.4 F 0.0
15:00 236.2 F 0.0
20:00 233.9 F 0.0
25:00 227.4 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

08/04/2022 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____