
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001415
10:07 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 5:13
STERILIZING 6:52
136.3 C 222kPa 6:52
136.4 C 222kPa 7:22
136.4 C 224kPa 7:52
136.3 C 221kPa 8:22
136.3 C 222kPa 8:52
136.4 C 223kPa 9:22
136.3 C 223kPa 9:52
136.2 C 221kPa 10:22
VENTING 10:22
AIR DRYING 11:23
CYCLE COMPLETE 11:33

Digital Signature #
0C356E79710282C5
