
Midmark M11 - v1.0.5
Total Cycles: 2251

Name: RITTER2
Sterilizer ID: V2313228
04/14/2023 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 75.2 F 0.0
4:00 84.0 F 0.0
6:00 104.7 F 0.0
8:00 152.4 F 1.5
10:00 170.3 F 4.8
12:00 196.8 F 8.4
14:00 217.6 F 12.3
16:00 219.2 F 5.7
18:00 230.5 F 9.8
20:00 236.1 F 10.6
22:00 239.3 F 11.2
24:00 236.4 F 9.2
26:00 249.4 F 15.4
28:00 260.5 F 21.7
30:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.4
2:00 271.8 F 29.5
3:00 272.0 F 29.2
4:00 272.1 F 29.8

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 205.5 F 0.0
10:00 209.4 F 0.0
15:00 209.6 F 0.0
20:00 206.5 F 0.0
25:00 215.0 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:46
CYCLE COMPLETE

04/14/2023 08:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____