

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 004386  
15:12 19/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:54  
STERILIZING 2:52  
136.3 C 222kPa 2:52  
136.4 C 225kPa 3:22  
136.1 C 223kPa 3:52  
136.2 C 223kPa 4:22  
136.2 C 222kPa 4:52  
136.0 C 221kPa 5:22  
136.4 C 220kPa 5:52  
136.3 C 225kPa 6:22  
136.3 C 225kPa 6:22  
VENTING 6:22  
AIR DRYING 7:20  
CYCLE COMPLETE 7:24

-----  
Digital Signature #  
36EFD81830736EB2  
-----