
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002303
8:57 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:16
STERILIZING 3:23
136.0 C 220kPa 3:23
136.4 C 221kPa 3:53
136.3 C 223kPa 4:23
136.3 C 222kPa 4:53
136.1 C 223kPa 5:23
136.4 C 222kPa 5:53
136.4 C 224kPa 6:23
136.4 C 224kPa 6:53
136.4 C 224kPa 6:53
VENTING 6:53
AIR DRYING 7:43
CYCLE COMPLETE 17:52

Digital Signature #
77B9579E12A379AA
