
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000730
11:46 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:27
STERILIZING 3:32
136.4 C 233kPa 3:32
136.0 C 230kPa 4:02
136.4 C 231kPa 4:32
136.4 C 233kPa 5:02
136.4 C 233kPa 5:32
136.3 C 233kPa 6:02
136.3 C 231kPa 6:32
136.3 C 233kPa 7:02
136.3 C 233kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 7:58

Digital Signature #
DAB769F580BFF30C
