

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007362  
7:15 13/05/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:05  
PRESSURIZING 4:05  
STERILIZING 5:12  
136.4 C 222kPa 5:12  
136.2 C 222kPa 5:42  
136.3 C 221kPa 6:12  
136.4 C 223kPa 6:42  
136.4 C 223kPa 7:12  
136.4 C 221kPa 7:42  
136.2 C 220kPa 8:12  
136.3 C 221kPa 8:42  
136.3 C 221kPa 8:42  
VENTING 8:42  
AIR DRYING 9:32  
CYCLE COMPLETE 9:45  
-----

Digital Signature #  
16CDDB3F316165D3  
-----