
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000793
12:25 16/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:59
PRESSURIZING 7:01
STERILIZING 8:44
136.0 C 220kPa 8:44
136.4 C 219kPa 9:14
136.3 C 221kPa 9:44
136.3 C 223kPa 10:14
136.4 C 223kPa 10:44
135.9 C 220kPa 11:14
136.4 C 223kPa 11:44
136.4 C 223kPa 12:14
136.4 C 223kPa 12:14
VENTING 12:14
AIR DRYING 13:24
CYCLE COMPLETE 32:38

Digital Signature #
FEEE7BF537D17783
