
Midmark M11 - v1.0.5
Total Cycles: 2379

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 197.3 F 4.2
4:00 210.9 F 10.1
6:00 223.4 F 6.5
8:00 237.5 F 12.9
10:00 240.3 F 11.8
12:00 240.9 F 11.4
14:00 255.3 F 18.8
16:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.8 F 29.9
2:00 271.8 F 29.2
3:00 271.8 F 29.7
4:00 272.6 F 29.8

Min 271.2 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 232.6 F 0.0
10:00 226.1 F 0.0
15:00 230.4 F 0.0
20:00 230.1 F 0.0
25:00 225.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 2:53
HEATING: 17:18
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:30
CYCLE COMPLETE

05/16/2023 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____