
Midmark M11 - v1.0.0
Total Cycles: 3547

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2023 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 175.9 F 1.3
4:00 186.3 F 4.6
6:00 211.1 F 9.0
8:00 220.8 F 8.0
10:00 230.5 F 9.3
12:00 239.6 F 9.6
14:00 242.4 F 11.8
16:00 239.7 F 10.1
18:00 252.9 F 16.6
20:00 264.1 F 23.2
22:00 271.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.9
2:00 271.5 F 28.1
3:00 272.1 F 28.8
4:00 272.2 F 28.7

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 234.9 F 0.0
10:00 238.6 F 0.0
15:00 231.1 F 0.0
20:00 226.6 F 0.0
25:00 228.2 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

07/19/2023 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____