
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001760
13:55 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:49
STERILIZING 3:58
136.2 C 222kPa 3:58
136.4 C 225kPa 4:28
136.4 C 223kPa 4:58
136.4 C 225kPa 5:28
136.4 C 223kPa 5:58
136.2 C 222kPa 6:28
136.3 C 225kPa 6:58
136.3 C 224kPa 7:28
136.3 C 224kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 8:37

Digital Signature #
BE31B36F899DABF1
