

-----  
Midmark M11 - v1.0.5  
Total Cycles: 406

Name: RITTER2  
Sterilizer ID: V2313228  
09/29/2021 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.0 F 0.0  
2:00 173.6 F 2.4  
4:00 187.0 F 6.6  
6:00 210.9 F 11.8  
8:00 224.0 F 7.2  
10:00 236.4 F 12.6  
12:00 238.2 F 10.6  
14:00 236.9 F 9.4  
16:00 251.3 F 16.2  
18:00 262.7 F 22.9  
20:00 272.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.0  
1:00 272.3 F 29.2  
2:00 272.4 F 29.5  
3:00 271.8 F 29.5  
4:00 271.7 F 28.9

Min 271.3 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 235.2 F 0.0  
10:00 221.7 F 0.0  
15:00 219.2 F 0.0  
20:00 222.9 F 0.0  
25:00 224.5 F 0.0  
30:00 221.8 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:23  
CYCLE COMPLETE

09/29/2021 11:03 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01