
Midmark M11 - v1.0.0
Total Cycles: 1598

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2022 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 139.7 F 1.3
4:00 155.8 F 4.7
6:00 177.7 F 8.6
8:00 197.9 F 13.1
10:00 213.9 F 4.9
12:00 227.8 F 9.4
14:00 236.6 F 10.4
16:00 239.2 F 10.8
18:00 236.5 F 8.4
20:00 249.9 F 15.0
22:00 261.6 F 21.6
24:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.0 F 28.5
2:00 272.6 F 28.6
3:00 271.5 F 28.0
4:00 272.4 F 28.9

Min 271.4 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 217.6 F 0.0
10:00 226.2 F 0.0
15:00 223.0 F 0.0
20:00 220.1 F 0.0
25:00 217.3 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

02/07/2022 10:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____