

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3717

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2024 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.4 F 0.0  
2:00 135.5 F 1.6  
4:00 146.3 F 4.3  
6:00 171.3 F 7.7  
8:00 192.7 F 11.6  
10:00 209.0 F 4.2  
12:00 217.2 F 8.4  
14:00 228.9 F 12.8  
16:00 232.5 F 9.2  
18:00 239.4 F 10.8  
20:00 244.6 F 13.1  
22:00 255.9 F 19.1  
24:00 266.1 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 271.9 F 29.9  
2:00 272.8 F 30.1  
3:00 272.4 F 29.8  
4:00 271.8 F 29.5

Min 271.1 F 29.0  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 207.3 F 0.0  
10:00 208.2 F 0.0  
15:00 206.2 F 0.0  
20:00 202.3 F 0.0  
25:00 200.3 F 0.0  
30:00 201.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:52  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:04:31  
CYCLE COMPLETE

06/19/2024 12:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_