
Midmark M11 - v1.0.5
Total Cycles: 4462

Name: RITTER2
Sterilizer ID: V2313228
03/26/2025 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 154.3 F 2.3
4:00 166.2 F 5.7
6:00 191.6 F 10.1
8:00 225.1 F 6.6
10:00 232.2 F 9.1
12:00 236.0 F 11.7
14:00 237.9 F 11.1
16:00 237.0 F 9.4
18:00 249.8 F 15.8
20:00 260.7 F 22.2
22:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.0
1:00 272.1 F 30.3
2:00 272.3 F 30.0
3:00 272.0 F 29.9
4:00 271.6 F 30.2

Min 271.5 F 29.9
Max 272.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 201.6 F 0.0
10:00 216.7 F 0.0
15:00 192.2 F 0.0
20:00 188.6 F 0.0
25:00 183.7 F 0.0
30:00 180.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

03/26/2025 05:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____