

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.2uS / 1.4ppm  
CYCLE NUMBER 000019  
16:13 13/01/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 4:12  
STERILIZING 6:08  
136.2 C 225kPa 6:08  
136.5 C 227kPa 6:38  
136.3 C 226kPa 7:08  
136.4 C 226kPa 7:38  
136.5 C 226kPa 8:08  
136.5 C 227kPa 8:39  
136.5 C 226kPa 9:09  
136.4 C 226kPa 9:38  
VENTING 9:38  
AIR DRYING 10:54  
CYCLE COMPLETE 12:54  
-----

Digital Signature #  
94A728CF9267E242  
-----