

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3533

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 10:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
2:00 196.5 F 1.6  
4:00 204.5 F 5.6  
6:00 216.6 F 10.2  
8:00 223.9 F 6.5  
10:00 233.1 F 9.5  
12:00 240.2 F 10.0  
14:00 242.7 F 12.2  
16:00 241.1 F 10.9  
18:00 253.7 F 17.2  
20:00 264.7 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.6 F 28.8  
2:00 271.4 F 28.1  
3:00 272.7 F 29.0  
4:00 271.9 F 28.4  
  
Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 228.2 F 0.0  
10:00 232.1 F 0.0  
15:00 224.0 F 0.0  
20:00 216.8 F 0.0  
25:00 216.5 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 21:59  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:11  
CYCLE COMPLETE

07/17/2023 11:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_