

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3899

Name: RITTER2  
Sterilizer ID: V2313228  
08/28/2024 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.8 F 0.0  
2:00 101.9 F 0.0  
4:00 106.3 F 0.3  
6:00 138.6 F 1.3  
8:00 174.8 F 4.2  
10:00 195.9 F 7.7  
12:00 209.0 F 11.6  
14:00 216.2 F 5.1  
16:00 221.0 F 9.4  
18:00 235.4 F 11.2  
20:00 236.5 F 10.9  
22:00 235.5 F 9.0  
24:00 248.7 F 15.2  
26:00 259.4 F 21.2  
28:00 269.1 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 271.9 F 29.9  
2:00 272.8 F 30.2  
3:00 272.6 F 29.9  
4:00 272.1 F 29.6

Min 271.0 F 29.1  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 221.8 F 0.0  
10:00 225.4 F 0.0  
15:00 211.7 F 0.0  
20:00 221.3 F 0.0  
25:00 230.9 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:17  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:07:45  
CYCLE COMPLETE

08/28/2024 12:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_