
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004505
9:37 06/10/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:39
STERILIZING 3:30
136.0 C 220kPa 3:30
136.0 C 222kPa 4:00
136.4 C 221kPa 4:30
136.3 C 224kPa 5:00
136.4 C 224kPa 5:30
136.4 C 223kPa 6:00
136.1 C 222kPa 6:30
136.1 C 221kPa 7:00
136.1 C 221kPa 7:00
VENTING 7:00
AIR DRYING 7:46
CYCLE COMPLETE 7:58

Digital Signature #
6DDBE092A098899F
