

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1945

Name: RITTER 1  
Sterilizer ID: V2178198  
05/18/2022 12:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.6 F 0.0  
2:00 144.2 F 2.5  
4:00 159.8 F 6.0  
6:00 187.3 F 10.5  
8:00 215.1 F 4.9  
10:00 223.2 F 8.6  
12:00 238.7 F 13.5  
14:00 236.4 F 9.9  
16:00 237.6 F 8.5  
18:00 247.5 F 14.0  
20:00 259.3 F 20.5  
22:00 270.3 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.2 F 28.8  
2:00 272.1 F 29.2  
3:00 271.8 F 28.5  
4:00 272.6 F 29.2

Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.3 F 0.0  
5:00 223.7 F 0.0  
10:00 228.4 F 0.0  
15:00 232.1 F 0.0  
20:00 241.0 F 0.0  
25:00 232.6 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

05/18/2022 01:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_