
Midmark M11 - v1.0.5
Total Cycles: 1765

Name: RITTER2
Sterilizer ID: V2313228
11/23/2022 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 80.2 F 0.0
4:00 98.9 F 0.0
6:00 147.4 F 0.0
8:00 176.6 F 0.2
10:00 182.2 F 0.5
12:00 200.5 F 3.5
14:00 217.8 F 6.8
16:00 228.7 F 10.5
18:00 233.9 F 8.4
20:00 238.5 F 11.1
22:00 236.1 F 8.9
24:00 240.5 F 10.1
26:00 248.0 F 14.3
28:00 259.8 F 20.9
30:00 270.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.2 F 29.0
2:00 272.3 F 29.0
3:00 272.6 F 29.2
4:00 272.8 F 29.5

Min 271.9 F 29.0
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 196.7 F 0.0
10:00 203.9 F 0.0
15:00 196.0 F 0.0
20:00 199.1 F 0.0
25:00 207.1 F 0.0
30:00 206.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:32
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

11/23/2022 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____