
Midmark M11 - v1.0.5
Total Cycles: 4072

Name: RITTER2
Sterilizer ID: V2313228
10/31/2024 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.3 F 0.0
2:00 92.8 F 0.0
4:00 98.1 F 0.0
6:00 136.8 F 0.3
8:00 165.7 F 1.7
10:00 186.6 F 4.6
12:00 205.3 F 7.9
14:00 212.4 F 11.6
16:00 225.0 F 6.4
18:00 231.5 F 11.0
20:00 233.5 F 7.8
22:00 243.6 F 13.5
24:00 242.5 F 12.3
26:00 254.3 F 18.2
28:00 264.6 F 24.5
30:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.7
1:00 272.4 F 29.7
2:00 272.8 F 30.0
3:00 272.5 F 30.3
4:00 271.7 F 29.8

Min 271.7 F 29.6
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 214.4 F 0.0
10:00 201.1 F 0.0
15:00 184.8 F 0.0
20:00 184.7 F 0.0
25:00 184.1 F 0.0
30:00 190.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

10/31/2024 03:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____