
Midmark M11 - v1.0.5
Total Cycles: 4169

Name: RITTER2
Sterilizer ID: V2313228
11/29/2024 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 142.3 F 0.1
4:00 161.7 F 1.0
6:00 189.3 F 4.4
8:00 209.5 F 8.2
10:00 221.6 F 12.7
12:00 228.5 F 8.1
14:00 239.6 F 13.1
16:00 239.2 F 11.2
18:00 238.6 F 10.3
20:00 251.9 F 16.8
22:00 263.4 F 23.5
24:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.6 F 30.0
2:00 271.8 F 29.7
3:00 272.6 F 30.0
4:00 271.9 F 29.5

Min 271.8 F 29.4
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 231.1 F 0.0
10:00 215.7 F 0.0
15:00 215.8 F 0.0
20:00 206.6 F 0.0
25:00 211.5 F 0.0
30:00 208.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:60
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:47
CYCLE COMPLETE

11/29/2024 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____