
Midmark M11 - v1.0.5
Total Cycles: 3983

Name: RITTER2
Sterilizer ID: V2313228
10/01/2024 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.7 F 0.0
2:00 131.0 F 0.3
4:00 140.8 F 2.8
6:00 167.9 F 6.5
8:00 187.6 F 10.1
10:00 205.0 F 7.4
12:00 216.5 F 7.7
14:00 227.5 F 12.0
16:00 231.9 F 8.6
18:00 241.4 F 12.3
20:00 243.2 F 12.4
22:00 254.4 F 18.2
24:00 264.4 F 24.4
26:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.9
1:00 272.8 F 30.0
2:00 272.6 F 30.4
3:00 272.0 F 29.9
4:00 272.9 F 30.1

Min 271.8 F 29.7
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 209.1 F 0.0
10:00 220.3 F 0.0
15:00 206.1 F 0.0
20:00 205.2 F 0.0
25:00 201.9 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:00
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:44
CYCLE COMPLETE

10/01/2024 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____