
Midmark M11 - v1.0.5
Total Cycles: 278

Name: RITTER2
Sterilizer ID: V2313228
08/10/2021 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.6 F 0.0
2:00 195.6 F 3.3
4:00 199.6 F 8.5
6:00 223.6 F 9.6
8:00 227.1 F 10.3
10:00 233.2 F 8.0
12:00 240.5 F 10.8
14:00 246.4 F 14.3
16:00 259.0 F 21.2
18:00 270.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 271.3 F 29.3
2:00 271.3 F 29.2
3:00 271.4 F 29.3
4:00 271.7 F 29.6

Min 271.0 F 29.2
Max 272.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 228.9 F 0.0
10:00 227.9 F 0.0
15:00 221.5 F 0.0
20:00 197.6 F 0.0
25:00 192.2 F 0.0
30:00 187.7 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 18:43
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

08/10/2021 02:31 PM
ACCEPTED: X
REJECTED:
Operator: U01