
Midmark M11 - v1.0.5
Total Cycles: 2560

Name: RITTER2
Sterilizer ID: V2313228
06/23/2023 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.0 F 0.0
2:00 196.6 F 4.0
4:00 209.4 F 9.5
6:00 223.2 F 7.2
8:00 235.6 F 11.6
10:00 236.1 F 9.7
12:00 237.1 F 8.9
14:00 250.3 F 15.8
16:00 262.5 F 22.9
18:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.7 F 29.7
2:00 272.1 F 29.0
3:00 271.5 F 29.1
4:00 272.7 F 29.8

Min 271.3 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 225.0 F 0.0
10:00 239.9 F 0.0
15:00 233.4 F 0.0
20:00 227.6 F 0.0
25:00 226.0 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

06/23/2023 02:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____