
Midmark M11 - v1.0.5
Total Cycles: 4156

Name: RITTER2
Sterilizer ID: V2313228
11/22/2024 04:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.9 F 0.0
2:00 150.0 F 0.7
4:00 178.0 F 4.8
6:00 205.2 F 10.0
8:00 220.4 F 5.6
10:00 232.1 F 11.2
12:00 233.7 F 8.8
14:00 237.3 F 9.6
16:00 248.1 F 15.0
18:00 261.2 F 22.3
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 272.5 F 29.4
2:00 272.4 F 29.7
3:00 273.0 F 30.0
4:00 272.1 F 29.8

Min 271.7 F 29.3
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 224.8 F 0.0
10:00 187.3 F 0.0
15:00 183.6 F 0.0
20:00 178.8 F 0.0
25:00 176.6 F 0.0
30:00 174.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

11/22/2024 05:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____