

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000123  
11:35 11/03/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:36  
PRESSURIZING 4:37  
STERILIZING 5:54  
136.0 C 220kPa 5:54  
136.2 C 223kPa 6:24  
136.4 C 224kPa 6:54  
136.4 C 224kPa 7:24  
136.2 C 224kPa 7:54  
136.2 C 223kPa 8:24  
136.4 C 224kPa 8:54  
136.1 C 221kPa 9:24  
136.1 C 221kPa 9:24  
VENTING 9:24  
AIR DRYING 10:26  
CYCLE COMPLETE 10:27

-----  
Digital Signature #  
DE053D0FCCD272F0  
-----